

Biomedical

### Introduction to Electrical Safety: Part II

### White Paper

Dennis J. McMahon, CBET-R

One of your principal duties as a medical equipment technician is to assure the safety of the many devices for which you are responsible. However, the standards for the design of this equipment are continually evolving, and it is important for you, the technician, to be current with the standards and how to test to them. This paper will give an overview of the intrinsic hazards of medical devices, electrical safety standards, and describe a testing procedure for the most commonly used standard in North America.

In Healthcare Technology Management (HTM), any discussion about electrical safety must begin with Ralph Nader's famous article in March of 1971, in which he alleged, "...about 1,200 people a year are electrocuted" in hospitals due to poor equipment design and inadequate or incompetent maintenance of that equipment. Nader's numbers may have been inflated, but his article set in motion a great deal of discussion about the safe design and maintenance of medical technology. This was a classic case of technology advancing much faster than the standards for its care. In the early 1970s, medical technology was beginning to evolve exponentially, but the standards for safe design and routine safety verification lagged behind the need. In the four decades since then, medical devices that diagnose and treat disease, or assist with treatment have advanced dramatically in complexity and number, and standards have been written and periodically amended to assure the equipment poses minimal risk both to the patient and the clinician. Patient safety should be paramount in the mind of the HTM technician. Our customers, the patients, are a vulnerable population who trust us to provide safe equipment for their care; our colleagues-physicians, nurses, and technologists-trust us to assure that the equipment is safe, as well as functional.

### A checklist for shock

Electrical safety is all about assuring that current flows only where it is designed to flow, and about minimizing the flow of unintended current. Unintended current is parasitic otherwise known as "leakage" current, that travels across the impedances that are inherent in any device. It cannot be totally eliminated. If there is sufficient leakage current, it poses the risk of shock, especially for patients whose clinical defenses are compromised.

#### A shock hazard exists if:

- A line-powered device is within the reach of a patient or user, **and**
- There are conductive parts available for the patient or user to contact, **and**
- The patient or user comes in contact with those parts, **and**
- Those parts become energized or ungrounded due to physical damage or component failure, **and**
- Enough current then passes through the patient or user to cause injury or death.

#### What can go wrong?

A few examples:

- The facility line voltage spikes above its nominal value by well over 10%.
- The connection to the protective earth wire in a poorly attached power plug opens up, causing the wire to be functionally disconnected.
- Years of constant flexing and stressing of the power cord, at either end, breaks most of the strands of the protective earth wire, severely increasing its resistance.
- A multi-parameter monitor falls to the floor and seems to be functional to the clinical users, who hurriedly restore it to use on a patient. But one of its circuit boards is partly dislodged or bent, and comes in contact with the grounded enclosure.
- Blood and saline solution, rich in electrolytes, spills from the bowl chamber of an autotransfusion system, dripping onto the line-energized power supply board.
- An ECG input connector is inadvertently shorted to the chassis after a repair, compromising its isolation from chassis ground.

Any of these, and countless other scenarios, can create excessive leakage current.



### Standards

Over the past four decades, a number of standards for the electrical safety of medical devices have been written. These documents are based on the best current design principles. Since components and design practices are always changing, the standards are revised periodically, typically every five years.

The benchmark standard worldwide is IEC 60601-1: Medical Electrical Equipment – General Requirements for Safety and Essential Performance, originally published in 1977 as IEC 601 by The International Electrotechnical Commission in Geneva. Primarily intended to be used for the design and production of medical devices as well as used by end-user facilities for periodic safety testing, it is currently in its third edition. Over the past thirty years it has been adopted, with slight variations, by a number of national agencies:

- British Standard EN 60601-1 (2006) in the U.K. and Europe.
- AS/NZS 3200.1 in Australia and New Zealand.
- JIS T 0601-1 in Japan.
- AAMI 60601-1 (2005) (R)2012 Association for the Advancement of Medical Instrumentation in the U.S.
- UL 60601-1 Underwriters Laboratories in the U.S. (Originally UL544)
- CAN/CSA C22.2 60601-1-08 by the Canadian Standards Association.

Standards have been published that parallel the intent of 60601-1, but are primarily for periodic safety verification by the clinical user facility, without the more rigorous "type testing" and production testing required of manufacturers. These include:

- NFPA-99 National Fire Protection Association in the U.S.
- ANSI/AAMI ES-1 (1993) in the U.S.
- DB 9801 in the U.K.
- VDE/ODE 751 in Germany and Austria
- AS/NZS 3551 in Australia and New Zealand

In May of 2007, the IEC published IEC 62353: Medical Electrical Equipment—Recurrent Test and Test After Repair, to harmonize with those standards intended for periodic in-house testing. Its tests reflect the higher level of protection designed into medical devices in the recent decade, and include earth bond testing, insulation testing, and three different ways to measure leakage current, depending on the type of device.

Additionally, in 2004 the IEC published its standard 61010-1: Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use. It includes devices used in the clinical laboratory, and so is pertinent to the HTM technician. If your routine maintenance includes laboratory equipment, you will need an analyzer that tests to this standard.

Ultimately, it is up to each facility's Clinical Engineering department to decide which testing standard it will apply, with deference to any governmental authority. (e.g. In the U.S., some state governments have adopted NFPA-99 as code in their jurisdiction, requiring its use by law.) Whatever standard is used, test results for each device must be recorded and retained.

### **Terminology and Definitions**

Comparing the major standards is made more challenging by differences in English language terminology between the North American and the European documents. The following terms are used in the text and associated schematics:

-- -- --

International:	North American:
L1	the 'Hot' or energized line
L2	the Neutral line
Earth	Ground
Enclosure or Case	Chassis
Mains Voltage	Line voltage (the facility
	wiring voltage)
Protective Earth	Ground Wire

#### **Applied Parts:**

Any conductive components that are intended to contact the patient in normal use. Examples: connectors for ECG or EEG electrodes, defibrillator pad connectors or paddles, active or return ESU electrodes.

#### Earth Leakage Current:

Current that flows in the Earth wire of a protectively earthed device.

# Enclosure Leakage Current (Touch or Chassis leakage):

Current that flows to earth from any exposed conductive part of the enclosure other than through the protective earth wire.

#### **Class I Device:**

Protection includes a Protective Earth wire.

#### **Class II Device:**

Class I protection, plus a second layer of enclosure insulation ('Double Insulated')

#### **Class III Device:**

Protection is based on the use of "Safe Extra-Low Voltage" (SELV): <25Vac or <60Vac. Generally battery-powered devices with no mains voltage.



## Patient Leakage Current (Lead to Ground Current):

Current that flows to earth through a patient connected to an applied part or parts.

#### **Patient Auxiliary Current:**

Current that flows between applied parts which serves no design purpose.

#### Isolation Leakage Current (Mains-on-Applied-Parts Current):

Current that flows to earth through the applied parts (connected as one) when momentarily connected to the mains voltage.

#### Single Fault Condition:

A condition in which a normal connection (Earth or L2) is open.

#### Type B equipment:

Complete protection against shock. Primarily external applications.

#### Type BF ('Body Floating") Equipment:

Applied parts to the patient's skin; input circuits are isolated.

#### Type CF ('Chest Floating') Equipment:

Applied parts including direct cardiac contact. >70M $\Omega$  from mains to applied parts.

#### **Testing Procedure**

Tests shown here apply to both the NFPA-99 and AAMI ES-1 standards, for line-powered devices that are used in the patient care vicinity, and portable (i.e. not 'fixed' devices which are physically secured to the building). North American terminology will be used.

The chart below is a condensed version of Table 1 of the ANSI/AAMI standard ES-1. "Fixed" equipment is permanently installed with an integrated ground (only removable with tools). In the Earth (Ground) Risk Current group, "Other" refers to devices that have no grounded accessible parts and no means of being grounded (this includes mobile X-ray equipment).

By separating-out a few exceptions, the table can be summarized for **portable devices** as follows:

#### **Enclosure Risk Current:**

Normal condition: 100 μA Single Fault Condition: 300 μA (except No Patient Contact devices: 500 μA)

#### **Patient Applied Parts Risk Current:**

Normal condition: 10 μA Single Fault Condition: Isolated devices: 50μA Non-Isolated devices: 100 μA

### Patient Isolated Risk Current: 50 µA

Consolidation of ANSI/AAMI ES-1, Table 1													
Enclosure Risk Current, µA													
	Isolated		Non-Isolated			Likely Pt Contact			No Pt Contact				
	Portable	Fixed	Portable	ible Fixed		Portable	Fixed		Portable	Fixed			
Normal	100	100	100	100		100	100		100	100			
Single Fault	300	5000	300	5000		300	5000		500	5000			
Patient Applied Part Current, µA													
		Isolated	Isolated		Non-Isolated			Patient Isolated					
Normal		10	10			10			n/a				
Single Fault		50	50			100			50				
Earth (Ground) Risk Current, µA													
		Coded	Coded			Other			Fixed				
Normal		500	500			2500			5000				
Single Fault		1000	1000			5000			1000				



## Earth (Ground) Risk Current for Corded Devices:

Normal condition: 500  $\mu A$  Single Fault condition: 1000  $\mu A$ 

## Earth (Ground) Risk Current for Other Devices:

Normal condition: 2500 µA Single Fault condition: 5000 µA

#### **The Electrical Safety Analyzer**

The choice of safety analyzer is driven primarily by the testing standard being used, and most analyzers are designed to test to a variety of standards. Regardless of which standard is used, the safety tests they have in common are:

- 1. Verification of the mains voltages
- 2. The resistance of the protective earth,
- 3. The current flow in the protective earth
- 4. Leakage current from the enclosure to ground,
- Leakage current from patient-contacting parts to ground,
- 6. Leakage current between patient-contacting parts, and
- 7. Leakage current from isolated patient-contacting parts when briefly connected to the mains voltage.

Additionally, some standards require all testing at 110% of line voltage, the Insulation Resistance, etc. The standard will determine those capabilities required of the analyzer. If the device under test has no applied parts, a more basic analyzer will do. Further considerations are the capability for data storage, automated test sequencing, and patient simulations.

Before testing, it is essential that the user understands the analyzer. Incompetent use of an analyzer can waste time, create erroneous test results, potentially damage the device under test, or even put the user at risk of shock. Additionally, many users do not take advantage of useful supplemental features of their analyzers, such as point-to-point tests or performance waveforms. Before testing, read the user manual, and practice testing on a functioning device.

#### **Test Sequence**

#### **1. Physical Inspection**

Some medical devices (especially portable devices) suffer a great deal of abuse between inspections, and clinical personnel may not always properly clean them or report damage, which could result in a safety hazard. Inspect the device for dents, cracks, or misalignments of enclosure sections. It is part of the inspection to take the time to clean the unit of the dirt, dust and dried liquid. Inspect the power cord, especially the plug. Inspect the patient interface components: cables, tubing, pads, cuffs, etc. Anything that is damaged or beyond cleaning for reasonably good appearance should be replaced.

#### 2. Configure the Safety Analyzer:

Select the standard that you will be using, then verify if the device to be tested is used in the patient care vicinity (defined in IEC 60601), and whether it is a Class I (line- powered) or Class II (double insulated) device. For some testing standards, you will need to know whether the device has Type B, Type BF, or Type CF isolation. For those tests which include reversing the polarity of the mains, the analyzer may offer a choice of time delays as a safeguard against damaging the unit under test.

#### 3. Mains Voltage:

Measure the voltage from Hot to Neutral, from Neutral to Ground, and from Hot to Ground. The voltage from Hot to Neutral and from Hot to Ground should be within +/- 10% of nominal. On 120v wiring, there should be no more or less than 3.6v from Neutral to Ground.



Example of a Mains Voltage measurement using the ESA609 Safety Analyzer.



#### 4. Ground Wire (Protective Earth) Resistance:

Several of the standards specify that the Ground Wire resistance should be measured first, before testing for any of the leakage currents. Since the values being measured here are very low, the resistance of the testing leads must first be zeroed, or nulled, before making a measurement.

It is a good practice (in fact specified in most standards) to stress the power cord at both connector ends while testing, in order to detect any incompetency of the ground wire continuity. The slightest amount of oxidation on the blades of the plug will also impact the resistance reading.



## 5. Ground Wire (Protective Earth) Leakage Current:

The analyzer is configured to measure the current in the Ground Wire.







## 6. Touch or Chassis (Enclosure) Leakage Current:

The analyzer measures current from the device's enclosure; usually, the "ground pin" is the best option. In the absence of a ground pin, you may connect to any conductive feature likely to be contacted by the patient or user, and which is continuous with the enclosure.





Example of a Chassis (Enclosure) Leakage Current Measurement using the ESA609 Electrical Safety Analyzer.

## 7. Patient (Applied Part to Ground) Leakage Current:

This measures the current between any and all conductive components which necessarily come in contact with the patient in normal use. Measurements are taken from each applied part, and from all applied parts connected together, Auxiliary ("Lead to Lead") Leakage Current:





Example of a Patient (Applied Part to Ground) Leakage Current measurement using the ESA612 Electrical Safety Analyzer.



#### 8. Auxiliary ("Lead to Lead") Leakage Current:

The analyzer is configured to determine the current between any two applied parts or any combination of applied part groups, user-selectable on the analyzer.



## 9. Isolation (Mains-on-Applied-Parts) Leakage Current:

In order to verify that the isolation of the applied parts is competent, this test exposes the applied parts, connected together, to the mains voltage created by an isolation transformer within the analyzer.



As members of the Clinical Engineering team, our goal is to assure the safety and reliability of the technology that is used for patient care. Electrical safety standards will continue to evolve and harmonize in the future, and an essential function of the technician is to be current with the standards, and apply them consistently.







### References

- Nader, R: "Danger in Our Hospitals: Ralph Nader's Most Shocking Expose'" Ladies Home Journal, 03/1971
- Eisner, L; Brown, R, & Modi, D: "Leakage Current Standards Simplified" MDDI Magazine, 07/2004
- 3. Associated Research, Inc.: "Three Ways to Simplify Medical Device Testing: Electrical Safety Testing per UL/IEC 60601 3rd Edition" 2012
- 4. Backes, John: "Safety Testing of Medical Devices: IEC 62353 Explained" European Medical Device Technology Magazine, 11/2007
- 5. ANSI/AAMI ES-1:1993 Safe Current Limits for Electromedical Apparatus The Association for the Advancement of Medical Instrumentation

To learn more about electrical safety or any of the analyzers discussed in this paper, please visit **www.flukebiomedical.com** 



#### About the author

Dennis J. McMahon, CBET-R, is a medical equipment technician with almost forty years of experience. After earning a degree in Chemistry in the late '60s, he worked as an anesthesia technician until 1977. He certified as a biomedical equipment technician in 1983, and worked for Virginia Mason Medical Center until 2011. He also volunteered in seven missions with Operation Smile between 2002 and 2011. Dennis has instructed in Healthcare Technology Management at North Seattle College since 2002, and has written a number of papers in the medical literature, and for Fluke Biomedical. He currently serves as the Education Chair for the Washington State Biomedical Association (WSBA).

#### Fluke Biomedical.

Trusted for the measurements that matter.

Fluke Biomedical 6045 Cochran Road Cleveland, OH 44139-3303 U.S.A.

For more information, contact us at: (800) 850-4608 or Fax (440) 349-2307 Email: sales@flukebiomedical.com Web access: www.flukebiomedical.com

©2015 Fluke Biomedical. Specifications subject to change without notice. Printed in U.S.A. 2/2015 6004694A EN

Modification of this document is not permitted without written permission from Fluke Corporation.